Product ID:ELECTRON VACUUM TUBE(S)

MSDS Date:03/15/1994

FSC:5960

NIIN:00-107-7585

MSDS Number: BXYCH === Responsible Party === Company Name:MPD, INC. Address:316 E 9TH ST. City:OWENSBORO

State:KY

ZIP:42301-2312

Country:US

Info Phone Num:502-685-6200

Emergency Phone Num:502-685-6200

CAGE:KO090 === Contractor Identification ===

Company Name: MPD INC. (FORMERLY GENERAL ELECTRIC CO)

Address:316 E 9TH STREET

City:OWENSBORO

State:KY

ZIP:42303-2312

Country:US

Phone:502-685-6200

CAGE:33173

Company Name:MPD, INC. Address:316 E 9TH ST. Box:City:OWENSBORO

State:KY

ZIP:42301-2312

Country:US

Phone:502-685-6200

CAGE:KO090

====== Composition/Information on Ingredients ========

Ingred Name:RHENIUM 187 (COMPOSES 2% OF THE HEATER WIRE FILAMENT) Other REC Limits:NONE RECOMMENDED

Ingred Name:TUNGSTEN (

CAS:7440-33-7 RTECS #:YO7175000 Other REC Limits:NONE RECOMMENDED OSHA PEL:5 MG/M3 ACGIH TLV:5 MG/M3
Ingred Name:RHENIUM 185 (NOT RADIOACTIVE-INGRED 2+ 3 COMPRISE THE BALANCE OF THE HEATER FILAMENT WIRE) Other REC Limits:NONE RECOMMENDED
=========== Hazards Identification ===========
Routes of Entry: Inhalation:NO Skin:YES Ingestion:NO Reports of Carcinogenicity:NTP:NO IAR C:NO OSHA:NO Health Hazards Acute and Chronic:ELECTRON TUBES OPERATE AT HIGH TEMPERATURES AND MAY CAUSE SKIN BURNS. Explanation of Carcinogenicity:THERE ARE NO INGREDIENTS ABOVE 0.1% WHICH ARE IDENTIFIED AS CARCINOGENS BY NTP,IARC OR OSHA. Effects of Overexposure:SKIN:THERMAL BURNS. Medical Cond Aggravated by Exposure:NONE SPECIFIED BY MANUFACTURER.
======================================
First Aid:NO EMERGENCY CARE IS ANTICIPATED.
======== F ire Fighting Measures ===============
Flash Point:NONE Extinguishing Media:THIS ITEM IS NOT FLAMMABLE. Fire Fighting Procedures:NO SPECIAL REQUIREMENTS. Unusual Fire/Explosion Hazard:THE ELECTRON TUBE IS UNDER VACUUM;IT MAY IMPLODE UNDER FIRE CONDITIONS.
========= Accidental Release Measures ==========
Spill Release Procedures:NOT REPORTED. Neutralizing Agent:NOT APPLICABLE.
======================================
Handling and Storag

INGREDIENTS 2 & 3 COMPRISE THE BALANCE OF THE COMPOSITION OF THE HEATER FILAMENT WIRE).

e Precautions:NO SPECIAL REQUIREMENTS
Other Precautions:NONE SPECIFIED BY MANUFACTURER.
====== Exposure Controls/Personal Protection ========
Respiratory Protection:NOT NEEDED IN NORMAL SERVICE. Ventilation:NOT NEEDED IN NORMAL SERVICE. Protective Gloves:COTTON Eye Protection:SAFETY GLASSES Other Protective Equipment:NONE Work Hygienic Practices:USE REASONABLE CARE IN HANDLING THIS PRODUCT.WASH HANDS AFTER USE AND BEFORE EATING. Supplemental Safety and Health NONE
======= Physical/Chemical Properties =============
HCC:A2 Solubility in Water:INSOLUBLE Appearance and Odor:SEALED GLASS TUBE,NO ODOR.
======================================
Stability Indicator/Materials to Avoid:YES NONE SPECIFIED BY MANUFACTURER. Stability Condition to Avoid:NONE SPECIFIED BY MANUFACTURER. Hazardous Decomposition Products:NONE SPECIFIED BY MANUFACTURER.
======== Disposal Considerations ============
Waste Disposal

Methods: DISPOSE OF IN ACCORDANCE WITH FEDERAL, STATE AND LOCAL REGULATIONS.

Disclaimer (provided with this information by the compiling agencies): This information is formulated for use by elements of the Department of Defense. The United States of America in no manner whatsoever, expressly or implied, warrants this information to be accurate and disclaims all liability for its use. Any person utilizing this document should seek competent professional advice to verify and assume res

nsibility for the suitability of this information to their articular situation.	